

TÜV Rheinland Energy GmbH
D-51101 Cologne



OPSIS AB
Mr. Svante Wallin
Box 244
244 02 Furulund
Sweden

svante.wallin@opsis.se

Mr. Martin Schneider
Phone +49 221 806-1614
Fax +49 221 806-1349
Mail martin.schneider@de.tuv.com
Cologne, 2019-10-18

Actual status of the type-approval testing for the OPSIS AB oxygen monitor O2000N

Dear Mr. Wallin,

With pleasure, we give you herewith an overview on the actual status of the suitability test of your measurement system O2000N for O₂.

The instrument has successfully passed the suitability test, including a laboratory and a three month field test. The system fulfills the requirements of the European standard EN 15267-3 (QAL1). The actual certificate is published under www.gal1.de.

Right now, the field test is going to be extended in order to achieve a longer maintenance interval.


A test report containing data of 1 year field testing has recently been send to the working group "Test reports" of the German committee BLAI to achieve a maintenance interval of 6 month. The next possible meeting of the working group "Test reports" of the German committee BLAI will take place in November 2019. If there will be positive overall assessment of the test report, the actual certificate will be extended with a maintenance interval of 6 month.

The test is still going on to finally achieve a maintenance interval of one year

Yours Sincerely

Environmental Protection

i.V.



Dipl.-Ing. Guido Baum

i. A.



Dipl.-Ing. Martin Schneider

TÜV Rheinland Energy GmbH
Am Grauen Stein
D-51105 Cologne

Phone +49 221 806-5200
Fax +49 221 806-1349
Mail energy@de.tuv.com
Web www.umwelt-tuv.de
www.eco-tuv.com

Managing Director
Dirk Fenske

District Court
Cologne HRB 56171