



सीएसआईआर-एनपीएल भारत प्रमाणन योजना
CSIR-NPL India Certification Scheme
(NPLI CS)

(सीएसआईआर-राष्ट्रीय भौतिक प्रयोगशाला: राष्ट्रीय मापन संस्थान, भारत)
डॉ. के.एस. कृष्णन मार्ग, नई दिल्ली-110012, भारत
(CSIR-National Physical Laboratory: National Metrology Institute of India)
Dr. K. S. Krishnan Marg, New Delhi-110012, India

VERIFICATION CERTIFICATE

Application No.: 25100906

Certificate No.: NPLI CS/ CEMS/ VC-2025-02

Date of issue: 17/10/2025

Valid up to: date of renewal
of the original certificate


This is to certify that the
Continuous Emission Monitoring System
Model: AR650 (IR)
Make: OPSIS

Manufactured by:
Opsis AB
P.O. Box 244 S-244 02 Furulund, Sweden

Issued to:
Nevco Engineers Pvt. Ltd.
90A-2nd Floor, Amritpuri B, East Of Kailash, Opposite Iskon Temple, New Delhi-110065

Previously Certified by:
MCERTS (Certificate no. Sira MC020011/08)

has been verified under CSIR-NPL India Certification Scheme (NPLI CS) and found suitable as per the defined guidelines of CPCB. This certificate is valid up to the next renewal date as mentioned in the previously issued certificate or any change notified by the CPCB in its Guideline, which will be applicable to this product or any part of this product. This certificate holder must have to submit the "Annual Surveillance Test" report to NICB for annual performance validation of this product.


Chairman Certification Committee
(NPL India Certification Scheme)

NPLI CS, CSIR-National Physical Laboratory
(National verification agency for certifying instruments and equipment for monitoring emissions and ambient air)
Dr. K. S. Krishnan Marg, New Delhi-110012, India

Fax : 91-11-45609310
NPLI CS : 91-11-45608287
General Enquiry : 91-11-45608650

E-mail : cfct.nplindia@csir.res.in
: ics.nplindia@csir.res.in
Website: www.nplindia.org

This NPLI CS document is valid only in its entirety. All the information within this document shall be considered for conditions of use. The content of this document must remain unaltered and may only be reproduced in its entirety. For any queries, please contact: ics.nplindia@csir.res.in



सीएसआईआर-एनपीएल भारत प्रमाणन योजना CSIR-NPL India Certification Scheme (NPLI CS)

(सीएसआईआर-राष्ट्रीय भौतिक प्रयोगशाला: राष्ट्रीय मापन संस्थान, भारत)
डॉ. के.एस. कृष्णन मार्ग, नई दिल्ली-110012, भारत
(CSIR-National Physical Laboratory: National Metrology Institute of India)
Dr. K. S. Krishnan Marg, New Delhi-110012, India

Application No.: 25100906

Certificate No.: NPLI CS/ CEMS/ VC-2025-02

Date of issue: 17/10/2025

Valid up to: date of renewal
of the original certificate

1. Details of Verified Parameter:

1.1 The product 650 (IR) verified after assessment of previously issued "Product Conformity Certificate" by MCERTS (Certificate no. Sira MC020011/08) for the following certification ranges:

| Sr. No. | Gases | Certified Range |
|---------|------------------|---------------------------|
| 1 | CO | 0 to 75 mg/m ³ |
| 2 | HCl | 0 to 15 mg/m ³ |
| 3 | H ₂ O | 0 to 30% Vol |
| 4 | HF | 0 to 5 mg/m ³ |

Table 1

1.2 The MCERTS certificate (Certificate no. Sira MC020011/08) issued for the Product 650 (IR) is only valid with the following components:

- 1.2.1 Receiver unit (Model RE062)
- 1.2.2 Transmitter unit (Model EM 062-A)
- 1.2.3 Control unit (analyser)

1.3 The certified operational conditions of the Product 650 (IR) are

- 1.3.1 Ambient Temperature Range: Stack components -30°C to +60°C
- 1.3.2 Control unit +5°C to +35°C
- 1.3.3 IP rating: IP20: Analyser, must be placed in a protected area
IP54: Duct mounted parts (transmitter & receiver unit)

This NPLI CS document is valid only in its entirety. All the information within this document shall be considered for conditions of use. The content of this document must remain unaltered and may only be reproduced in its entirety. For any queries, please contact: ics.nplindia@csir.res.in



सीएसआईआर-एनपीएल भारत प्रमाणन योजना CSIR-NPL India Certification Scheme (NPLI CS)

(सीएसआईआर-राष्ट्रीय भौतिक प्रयोगशाला: राष्ट्रीय मापन संस्थान, भारत)
डॉ. के.एस. कृष्णन मार्ग, नई दिल्ली-110012, भारत
(CSIR-National Physical Laboratory: National Metrology Institute of India)
Dr. K. S. Krishnan Marg, New Delhi-110012, India

Application No.: 25100906

Certificate No.: NPLI CS/ CEMS/ VC-2025-02

Date of issue: 17/10/2025

Valid up to: date of renewal
of the original certificate

1.4 The certified performance of Product 650 (IR)

| Sr. No. | Test Parameters | CO | HCl | H ₂ O | HF |
|---------|--|---------|---------|------------------|---------|
| 1 | Response time | < 120 s | < 120 s | < 120 s | < 120 s |
| 2 | Repeatability standard deviation at zero point | 0.8 | 1.8 | 0.1 | 0.8 |
| 3 | Repeatability standard deviation at reference point | 0.4 | 1.3 | 0.2 | 0.9 |
| 4 | Lack-of-fit | 0.93 | 2.00 | 0.67 | -1.80 |
| 5 | Influence of ambient temperature zero point | -0.40 | 0.20 | 0.00 | -1.20 |
| 6 | Influence of ambient temperature reference point | -1.20 | 0.20 | 0.03 | 4.00 |
| 7 | Influence of sample gas pressure | -0.53 | -0.89 | -0.33 | -0.47 |
| 8 | Influence of voltage variations 190 to 250V | 0.50 | -0.90 | -0.10 | -0.80 |
| 9 | Cross-sensitivity at zero with interferents: O ₂ , H ₂ O, CO, CO ₂ , CH ₄ , N ₂ O, NO, NO ₂ , NH ₃ , SO ₂ , HCl | -0.44 | 0.00 | -0.67 | 0.00 |
| 10 | Cross-sensitivity at reference with interferents: O ₂ , H ₂ O, CO, CO ₂ , CH ₄ , N ₂ O, NO, NO ₂ , NH ₃ , SO ₂ , HCl | 0.84 | 0.94 | 0.57 | 0.00 |

1.5 The instrument exhibited some moderate resonances. Some resonances caused the light source to go out. The effect was only temporary, and the system functioned correctly once restored.

1.6 All deviations below 0.5% are considered to be negligible and not reported.

1.7 Based on the field calibration function test and the laboratory lack-of-fit test. The lack of fit in the field must be verified during every check of the installation of the CEM.

This NPLI CS document is valid only in its entirety. All the information within this document shall be considered for conditions of use. The content of this document must remain unaltered and may only be reproduced in its entirety. For any queries, please contact: ics.nplindia@csir.res.in



सीएसआईआर-एनपीएल भारत प्रमाणन योजना CSIR-NPL India Certification Scheme (NPLI CS)

(सीएसआईआर-राष्ट्रीय भौतिक प्रयोगशाला: राष्ट्रीय मापन संस्थान, भारत)
डॉ. के.एस. कृष्णन मार्ग, नई दिल्ली-110012, भारत
(CSIR-National Physical Laboratory: National Metrology Institute of India)
Dr. K. S. Krishnan Marg, New Delhi-110012, India

Application No.: 25100906

Certificate No.: NPLI CS/ CEMS/ VC-2025-02

Date of issue: 17/10/2025

Valid up to: date of renewal
of the original certificate

2. Acceptability of the product for use in Indian Conditions:

- 2.1 The measuring instrument AR 650 (IR) complies with the Indian environmental conditions.
- 2.2 The AR 650 (IR) system is a cross-stack flue gas measurement system.
- 2.3 The uses of AR650 (IR), depending on the parameter-specific Emission Standards for Indian industries as described in Table 8 of the 1st Revised Guidelines for Continuous Emission Monitoring Systems, August 2018 published by CPCB.
- 2.4 The use of AR650 (IR) technology for measurement of flue gases, depending on "Technical Selection & Suitability for Gaseous CEMS" as described in Table 5 of the 1st Revised Guidelines for Continuous Emission Monitoring Systems, August 2018 published by CPCB.

3. General Note:

After assessment of the MCERTS (Certificate no. Sira MC020011/08), the Certification Committee (NPLI CS) reconfirmed that the product passed the acceptability criteria for Indian uses and is eligible for this Verification Certificate.

It is the responsibility of M/s Nevco Engineers Pvt. Ltd., 90A-2nd Floor, Amritpuri B, East Of Kailash, Opposite Iskon Temple, New Delhi-110065 (On behalf of Opsis AB, P.O. Box 244 S-244 02 Furulund, Sweden) to inform CSIR-NPL about the manufacturing modifications in the model no. AR650 (IR).

It is also the responsibility of M/s Nevco Engineers Pvt. Ltd., 90A-2nd Floor, Amritpuri B, East Of Kailash, Opposite Iskon Temple, New Delhi-110065 (On behalf of Opsis AB, P.O. Box 244 S-244 02 Furulund, Sweden) to provide the list of installation of this equipment (with geographic coordinates) to CSIR-NPL for surveillance.

This Verification Certificate is only valid for the **Model No. AR650 (IR)**.

The validity of this Verification Certificate is bound to the renewal date of the original Certificate no. Sira MC020011/08.

This NPLI CS document is valid only in its entirety. All the information within this document shall be considered for conditions of use. The content of this document must remain unaltered and may only be reproduced in its entirety. For any queries, please contact: ics.nplindia@csir.res.in