

**Certificate on Product Conformity (QAL1)
for AR500 for NO₂, SO₂ and O₃**

**No. of Certificate: 0000025927_02, TÜV Rheinland Energie und
Umwelt GmbH, Cologne, Germany, 2 February 2015**

CERTIFICATE

of Product Conformity (QAL1)

Certificate No. : 0000025927_02

Certified AMS: AR500 with ER120 for NO₂, SO₂ and O₃

Manufacturer: Opsis AB
Skytteskogsvägen 16
244 02 Furulund
Sweden

Test Institute: TÜV Rheinland Energie und Umwelt GmbH

**This is to certify that the AMS has been tested
and found to comply with:**

**EN 14211: 2005, EN 14212: 2005, EN 14625: 2005,
EN 15267-1: 2009 and EN 15267-2: 2009**

Certification is awarded in respect of the conditions stated in this certificate
(see also the following pages).




Suitability Tested
Complying with
2008/50/EC
EN 15267
Regular
Surveillance

www.tuv.com
ID 0000025927

Publication in the German Federal Gazette
(BAnz.) of 2 March 2012


German Federal Environment Agency
Dessau, 2 February 2015



i. A. Dr. Marcel Langner

This certificate will expire on:
11 February 2020

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 30 January 2015



ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Accreditation according to EN ISO/IEC 17025 and certified according to ISO 9001:2008.

Certificate:
0000025927_02 / 2 February 2015

Test report: 936/21211350/B of 7 October 2011
Initial certification: 12 February 2010
Certificate: renewal (previous certificate 0000025927_01 of 16 March 2012
valid until 11 February 2015)
Date of expiry: 11 February 2020
Publication: BAnz. 2 March 2012, no. 36, p. 920, chapter IV, no. 2.1

Approved application

The certified AMS is suitable for the continuous monitoring of concentrations of nitrogen dioxide, sulphur dioxide, and ozone in ambient air (stationary operation).

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test.

The AMS is approved for the temperature range of +5 °C to +40 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21211350/B dated 7 October 2011 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette (BAnz. 2 March 2012, no. 36, p. 920, chapter IV, No. 2.1, UBA announcement of 23 February 2012)
- publication in the German Federal Gazette (BAnz AT 5 March 2013 B10, chapter V, notice 11, UBA announcement of 12 February 2013)

AMS designation:

AR500 with ER120 for NO₂, SO₂ and O₃

Manufacturer:

Opsis AB, Furulund, Sweden

Field of application:

For the continuous monitoring of concentrations of nitrogen dioxide, sulphur dioxide and ozone in ambient air (stationary operation)

Measuring ranges during the performance test:

Component	Certification range	supplementary measurement ranges	Unit
NO ₂	0 - 400	0 - 1800	µg/m ³
SO ₂	0 - 700	0 - 1000	µg/m ³
O ₃	0 - 360	0 - 500	µg/m ³

Software version:

7.21

Restrictions:

None

Notes:

1. The measuring path length during performance testing was 320 m.
2. The maintenance interval is four weeks.
3. Equivalence with the reference measurement methods according to the guideline „Demonstration of Equivalence of Ambient Air Monitoring Methods“ has been demonstrated for the components NO₂, SO₂ and O₃.
4. Functional tests by external sample gas feeding are possible.
5. The report on the performance test is available online at www.qal1.de.
6. Supplementary testing (Demonstration of Equivalence for the component SO₂ according the guideline “Demonstration of Equivalence of Ambient Air Monitoring Methods”) as regards Federal Environment Agency notices of 25 January 2010 (BAnz. p. 552, chapter III no. 1.1).

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report no.: 936/21211350/B dated 7 October 2011

11 Notification as regards Federal Environment Agency notices of 23 February 2012 (Federal Gazette (BAnz.) p. 920, chapter IV, no. 2.1)

The AR500 measuring system with ER120 for NO₂, SO₂ and O₃ manufactured by Opsis AB can also be operated with the transmitter/receiver units ER 110 and ER150.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 10 October 2012

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The measurement system AR500 operates on the basis of the Differential Optical Absorption Spectroscopy (DOAS). The DOAS measuring principle uses the characteristic radiation absorption by gaseous components for quantification of the respective concentrations.

The DOAS monitor AR500 with ER120 consists of a combined transmitter-receiver unit, a reflector unit and an analyser. The absorbed light is transferred from the transmitter-receiver unit to the analyser via fibre optic cable.

Combined transmitter-receiver Unit ER 120

The combined transmitter-receiver unit ER120 comprises the optical components, the xenon light-source and the power supply PS150 for igniting the xenon light-source.

The used high-pressure xenon lamp is a point light source. The light is generated by ignition of ultra pure xenon gas at a pressure of approx. 30 bar. The lamp is powered by a stabilised D.C. voltage source and requires a short high-voltage ignition pulse.

The radiation of the lamp includes the ultraviolet, the visible, and the infrared range. The wavelengths are continuously distributed over the entire spectrum, with the exception of some peaks in the near-infrared range.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energie und Umwelt GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet: qal1.de.

Certification of the AR500 with ER120 for NO₂, SO₂ and O₃ is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267:

Certificate No. 0000025927: 10 March 2010

Validity of the certificate: 11 February 2015

Test report: 936/21211350/A of 26 October 2009,
TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz. 12 February 2010, no. 24, p. 552, chapter III, no. 1.1
UBA announcement of 25 January 2010

Supplementary testing according to EN 15267:

Certificate No. 0000025927_01: 16 March 2012

Validity of the certificate: 11 February 2015

Test report: 936/21211350/B of 7 October 2011,
TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz. 2 March 2012, no. 36, p. 920, chapter IV, no. 2.1
UBA announcement of 23 February 2012

Notifications

Statement of TÜV Rheinland Energie und Umwelt GmbH of 10 October 2012
Publication: BAnz AT 5 March 2013 B10, chapter V, notification 11 (additional transmitter/receiver)
UBA announcement of 12 February 2013

Renewal of the certificate

Certificate No. 0000025927_02 2 February 2015

Expiry date of the certificate: 11 February 2020

Table 1: Total expanded uncertainty with the results of the laboratory test according to EN 14211 (component NO₂) for system 1329

Device: AR500		Serial No: 1329		1h-limit value: 104,6		nmol/mol	
Component: NO ₂		1h-limit value:		104,6		nmol/mol	
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability at zero	≤ 1,0 nmol/mol	0,000	U _{r,z} 0,00	0,0000		
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	2,000	U _{r,th} 0,04	0,0015		
3	"lack of fit"	≤ 4,0% of measured value	0,800	U _{l,th} 0,48	0,2334		
4	Sensitivity coefficient of sample gas pressure	≤ 8,0 nmol/mol/kPa	0,000	U _{gp} 0,00	0,0000		
5	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/K	0,026	U _{gt} 0,04	0,0016		
6	Sensitivity coefficient of surrounding temperature	≤ 3,0 nmol/mol/K	-0,050	U _{st} -0,06	0,0036		
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,021	U _v -0,07	0,0046		
8a	H2O with concentration 21 nmol/mol	≤ 5,0 nmol/mol	0,000	U _{H2O} 0,00	0,0000		
8b	CO2 with concentration 500 µmol/mol	≤ 5,0 nmol/mol	0,001	U _{int,pos}	0,2304		
8c	O3 with concentration 200 nmol/mol	≤ 2,0 nmol/mol	0,002	or			
8d	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,002	U _{int,neg}			
9	Averaging effect	≤ 7,0% of measured value	-0,600	U _{av} -0,36	0,1313		
18	Difference sample/calibration port	≤ 1,0%	0,000	U _{asc} 0,00	0,0000		
21	Converter efficiency	≥ 98%	100,000	U _{CE} 0,00	0,0000		
22	Increase of NO2 concentration due to residence time	≤ 4,0 nmol/mol	0,000	U _{ctr} 0,00	0,0000		
23	Uncertainty calibration gas	≤ 3,0%	2,000	U _{cg} 1,05	1,0941		
				combined standard uncertainty	U _c	1,3046	nmol/mol
				expanded uncertainty	U _c	2,6092	nmol/mol
				expanded uncertainty actual	U _{c,rel}	2,49	%
				expanded uncertainty required	U _{req,rel}	15	%

Table 2: Total expanded uncertainty with the results of the laboratory test and field test according to EN 14211 (component NO₂) for system 1329

Device: AR500		Serial No: 1329		1329		nmol/mol	
Component: NO ₂		1h-limit value:		104,6			
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability at zero	1,0 nmol/mol	0,000	0,00	0,0000		
2	Repeatability at concentration ct	3,0 nmol/mol	2,000	not respected because, $u_{r,1h} = 0,075 < u_{r,f}$	-		
3	"lack of fit"	4,0% of measured value	0,800	0,48	0,2334		
4	Sensitivity coefficient of sample gas pressure	8,0 nmol/mol/kPa	0,000	0,00	0,0000		
5	Sensitivity coefficient of sample gas temperature	3,0 nmol/mol/K	0,026	0,04	0,0016		
6	Sensitivity coefficient of surrounding temperature	3,0 nmol/mol/K	-0,050	-0,06	0,0036		
7	Sensitivity coefficient of electrical voltage	0,30 nmol/mol/V	-0,021	-0,07	0,0046		
8a	H2O with concentration 21 nmol/mol	5,0 nmol/mol	0,000	0,00	0,0000		
8b	CO2 with concentration 500 µmol/mol	5,0 nmol/mol	0,001	0,48	0,2304		
8c	O3 with concentration 200 nmol/mol	2,0 nmol/mol	0,002				
8d	NH3 with concentration 200 nmol/mol	5,0 nmol/mol	0,002				
9	Averaging effect	7,0% of measured value	-0,600	-0,36	0,1313		
10	Reproducibility under field conditions	5,0% of the average of 3 Mon.	4,720	4,94	24,3752		
11	Long term drift at zero level	5,0 nmol/mol	-1,420	-0,82	0,6721		
12	Long term drift at span level	5,0% of max. of certification range	0,430	0,26	0,0674		
18	Difference sample/calibration port	1,0%	0,000	0,00	0,0000		
21	Converter efficiency	98%	100,000	0,00	0,0000		
22	Increase of NO2 concentration due to residence time	4,0 nmol/mol	0,000	0,00	0,0000		
23	Uncertainty calibration gas	3,0%	2,000	1,05	1,0941		
combined standard uncertainty				u_c	7,1546	nmol/mol	
expanded uncertainty				U_c	14,3093	nmol/mol	
expanded uncertainty actual				$U_{c,rel}$	13,68	%	
expanded uncertainty required				$U_{req,rel}$	15	%	

Table 3: Total expanded uncertainty with the results of the laboratory test according to EN 14211 (component NO₂) for system 1330

Device: AR500		Serial No: 1330		nmol/mol	
Component: NO ₂		1h-limit value:		104,6	
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	U _{r,z}	0,0003
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,900	U _{r,th}	0,0004
3	"lack of fit"	≤ 4,0% of measured value	0,000	U _{l,th}	0,1313
4	Sensitivity coefficient of sample gas pressure	≤ 8,0 nmol/mol/kPa	0,000	U _{sp}	0,0000
5	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/K	0,000	U _{gt}	0,0025
6	Sensitivity coefficient of surrounding temperature	≤ 3,0 nmol/mol/K	0,000	U _{st}	0,0000
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,000	U _v	0,0553
8a	H2O with concentration 21 nmol/mol	≤ 5,0 nmol/mol	0,000	U _{H2O}	0,0000
8b	CO2 with concentration 500 µmol/mol	≤ 5,0 nmol/mol	0,000	U _{int,pos}	0,1764
8c	O3 with concentration 200 nmol/mol	≤ 2,0 nmol/mol	0,000	or	
8d	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,000	U _{int,neg}	0,0328
9	Averaging effect	≤ 7,0% of measured value	0,000	U _{av}	
18	Difference sample/calibration port	≤ 1,0%	0,000	U _{asc}	0,0000
21	Converter efficiency	≥ 98%	100,000	U _{CE}	0,0000
22	Increase of NO2 concentration due to residence time	≤ 4,0 nmol/mol	0,000	U _{ctr}	0,0000
23	Uncertainty calibration gas	≤ 3,0%	2,000	U _{cg}	1,0941
combined standard uncertainty				U _c	1,2222
expanded uncertainty				U _e	2,4445
expanded uncertainty actual				U _{e,rel}	2,34
expanded uncertainty required				U _{req,rel}	15

Table 4: Total expanded uncertainty with the results of the laboratory test and field test according to EN 14211 (component NO₂) for system 1330

Device:		AR500		Serial No:		1330	
Component:		NO ₂		1h-limit value:		104,6	
nmol/mol							
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	u _{r,z}	0,02	0,0003	
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,900	u _{r,h}	not respected because, u _{r,h} = 0,034 < u _{r,f}	-	
3	"lack of fit"	≤ 4,0% of measured value	0,600	u _{l,h}	0,36	0,1313	
4	Sensitivity coefficient of sample gas pressure	≤ 8,0 nmol/mol/kPa	0,000	u _{pp}	0,00	0,0000	
5	Sensitivity coefficient of sample gas temperature	≤ 3,0 nmol/mol/K	-0,032	u _{pt}	-0,05	0,0025	
6	Sensitivity coefficient of surrounding temperature	≤ 3,0 nmol/mol/K	0,000	u _{st}	0,00	0,0000	
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,073	u _v	0,24	0,0553	
8a	H2O with concentration 21 nmol/mol	≤ 5,0 nmol/mol	0,000	u _{H2O}	0,00	0,0000	
8b	CO2 with concentration 500 µmol/mol	≤ 5,0 nmol/mol	0,001	u _{CO2,pos}	0,42	0,1764	
8c	O3 with concentration 200 nmol/mol	≤ 2,0 nmol/mol	0,002	or			
8d	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,000	u _{NH3,neg}			
9	Averaging effect	≤ 7,0% of measured value	-0,300	u _{av}	-0,18	0,0328	
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,720	u _{r,f}	4,94	24,3752	
11	Long term drift at zero level	≤ 5,0 nmol/mol	1,620	u _{l,z}	0,94	0,8748	
12	Long term drift at span level	≤ 5,0% of max. of certification range	0,500	u _{l,h}	0,30	0,0912	
18	Difference sample/calibration port	≤ 1,0%	0,000	u _{Dsc}	0,00	0,0000	
21	Converter efficiency	≤ 0,98	100,000	u _{CE}	0,00	0,0000	
22	Increase of NO2 concentration due to residence time	≤ 4,0 nmol/mol	0,000	u _{ctr}	0,00	0,0000	
23	Uncertainty calibration gas	≤ 3,0%	2,000	u _{cg}	1,05	1,0941	
combined standard uncertainty				u _c	7,1561 nmol/mol		
expanded uncertainty				U _c	14,3121 nmol/mol		
expanded uncertainty actual				U _{c,rel}	13,68 %		
expanded uncertainty required				U _{req,rel}	15 %		

Table 5: Total expanded uncertainty with the results of the laboratory test according to EN 14212 (component SO₂) for system 1329

Device: AR500		Serial-No.: Gerät 1 (1329)		1h-limit value:		nmol/mol	
Component: SO ₂		1h-limit value:		1h-limit value:		1h-limit value:	
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	u _{r,z}	0,0003		
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	u _{r,lv}	0,0003		
3	"lack of fit"	≤ 4,0% of measured value	1,600	u _{lv}	1,4868		
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/kPa	0,000	u _{sp}	0,0000		
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,071	u _{tr}	0,2908		
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,030	u _{st}	0,0523		
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,010	u _v	0,0103		
8a	H2O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	u _{H2O}	0,0000		
8b	H2S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,409	u _{H2S,pos}	0,1600		
8c	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,406	or			
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	-0,604	or			
8e	NO2 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,404	or			
8f	m-Xylol with concentration 1 µmol/mol	≤ 10 nmol/mol	1,421	u _{int,neg}	0,0058		
9	Averaging effect	≤ 7,0% of measured value	-0,100	u _{av}			
18	Difference sample/calibration port	≤ 1,0%	0,000	u _{pac}	0,0000		
23	Uncertainty calibration gas	≤ 3,0%	2,000	ucg	1,7424		
		combined standard uncertainty	u _c	u _c	1,9363	nmol/mol	
		expanded uncertainty	U _c	U _c	3,8726	nmol/mol	
		expanded uncertainty actual	U _{c,rel}	U _{c,rel}	2,93	%	
		expanded uncertainty required	U _{req,rel}	U _{req,rel}	15	%	

Table 6: Total expanded uncertainty with the results of the laboratory test and field test according to EN 14212 (component SO₂) for system 1329

Device: AR500		Serial-No.: 132		Gerät 1 (1329)		nmol/mol	
Component: SO ₂		1h-limit value:		Square of uncertainty		nmol/mol	
No.	Performance characteristic	Criterion	Result	Uncertainty	U _z	U _z	U _z
1	Repeatability at zero	≤ 1,0 nmol/mol	0,100	0,02	0,0003		
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	not respected because: $u_{r,IV} = 0,01 < u_{r,f}$	-		
3	"lack of fit"	≤ 4,0% of measured value	1,600	1,22	1,4868		
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/kPa	0,000	0,00	0,0000		
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,071	0,54	0,2908		
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,030	-0,23	0,0523		
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	-0,010	-0,10	0,0103		
8a	H2O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	0,00	0,0000		
8b	H2S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,409	0,40	0,1600		
8c	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,406				
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	-0,604				
8e	NO2 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	-0,404				
8f	m-Xylol with concentration 1 µmol/mol	≤ 10 nmol/mol	1,421				
9	Averaging effect	≤ 7,0% of measured value	-0,100	-0,08	0,0058		
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,830	6,38	40,6483		
11	Long term drift at zero level	≤ 5,0 nmol/mol	-0,920	-0,53	0,2821		
12	Long term drift at span level	≤ 5,0% of max. of certification range	1,490	1,14	1,2894		
18	Differenz Proben-/Kalibrierungsgang	≤ 1,0%	0,000	0,00	0,0000		
23	Unsicherheit Prüfgas	≤ 3,0%	2,000	1,32	1,7424		
combined standard uncertainty						u _c	6,7800
expanded uncertainty						U _c	13,5600
expanded uncertainty actual						U _{c,rel}	10,27
expanded uncertainty required						U _{req,rel}	15

Table 7: Total expanded uncertainty with the results of the laboratory test according to EN 14212 (component SO₂) for system 1330

Device: AR500		Serial-No.: 132		Gerät 2 (1330)		nmol/mol	
Component: SO ₂		1h-limit value:		Square of uncertainty			
No.	Performance characteristic	Criterion	Result	Uncertainty			
1	Repeatability at zero	≤ 1,0 nmol/mol	0,000	U _{r,z}	0,00	0,0000	
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	U _{r,lv}	0,02	0,0003	
3	"lack of fit"	≤ 4,0% of measured value	1,400	U _{lv}	1,07	1,1384	
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/kPa	0,000	U _{sp}	0,00	0,0000	
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,011	U _{gt}	0,08	0,0070	
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,060	U _{st}	-0,46	0,2091	
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,010	U _v	0,10	0,0103	
8a	H2O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	U _{H2O}	0,00	0,0000	
8b	H2S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,503	U _{in,pos}			
8c	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,203				
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	0,202	or	1,23	1,5129	
8e	NO2 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,401				
8f	m-Xylol with concentration 1 µmol/mol	≤ 10 nmol/mol	0,809	U _{in,neg}			
9	Averaging effect	≤ 7,0% of measured value	0,100	U _{av}	0,08	0,0058	
18	Difference sample/calibration port	≤ 1,0%	0,000	U _{osc}	0,00	0,0000	
23	Uncertainty calibration gas	≤ 3,0%	2,000	0	1,32	1,7424	
				combined standard uncertainty	U _c	2,1509	nmol/mol
				expanded standard uncertainty	U _e	4,3017	nmol/mol
				expanded uncertainty actual	U _{e,rel}	3,26	%
				expanded uncertainty required	U _{req,rel}	15	%

Table 8: Total expanded uncertainty with the results of the laboratory test and field test according to EN 14212 (component SO₂) for system 1330

Device: AR500		Serial-No.: 132		Gerät 2 (1330)		nmol/mol	
Component: SO ₂		1h-limit value:		Square of uncertainty			
No.	Performance characteristic	Criterion	Result	Uncertainty	U _{r,z}	U _{r,z}	
1	Repeatability at zero	≤ 1,0 nmol/mol	0,000	0,00	0,0000		
2	Repeatability at concentration ct	≤ 3,0 nmol/mol	0,100	not respected because, $u_{r,z} = 0,01 < u_{r,f}$	-		
3	"lack of fit"	≤ 4,0% of measured value	1,400	1,07	1,1384		
4	Sensitivity coefficient of sample gas pressure	≤ 3,0 nmol/mol/kPa	0,000	0,00	0,0000		
5	Sensitivity coefficient of sample gas temperature	≤ 1,0 nmol/mol/K	0,011	0,08	0,0070		
6	Sensitivity coefficient of surrounding temperature	≤ 1,0 nmol/mol/K	-0,060	-0,46	0,2091		
7	Sensitivity coefficient of electrical voltage	≤ 0,30 nmol/mol/V	0,010	0,10	0,0103		
8a	H2O with concentration 21 nmol/mol	≤ 10 nmol/mol	0,000	0,00	0,0000		
8b	H2S with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,503				
8c	NH3 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,203				
8d	NO with concentration 500 nmol/mol	≤ 5,0 nmol/mol	0,202				
8e	NO2 with concentration 200 nmol/mol	≤ 5,0 nmol/mol	0,401				
8f	m-Xylol with concentration 1 µmol/mol	≤ 10 nmol/mol	0,809				
9	Averaging effect	≤ 7,0% of measured value	0,100	0,08	0,0058		
10	Reproducibility under field conditions	≤ 5,0% of the average of 3 Mon.	4,830	6,38	40,6483		
11	Long term drift at zero level	≤ 5,0 nmol/mol	1,160	0,67	0,4485		
12	Long term drift at span level	≤ 5,0% of max. of certification range	-2,070	-1,58	2,4887		
18	Differenz Proben-/Kalibrierungsgang	≤ 1,0%	0,000	0,00	0,0000		
23	Unsicherheit Prüfgas	≤ 3,0%	2,000	1,32	1,7424		
		combined standard uncertainty	expanded uncertainty	U _c	6,9434	nmol/mol	
		expanded uncertainty actual	U _c		13,8869	nmol/mol	
		expanded uncertainty required	U _{c,rel}		10,52	%	
				U _{req,rel}	15	%	

Table 9: Total expanded uncertainty with the results of the laboratory test according to EN 14625 (component O₃) for system 1329

Device: AR500		Serial No.		Gerät 1 (1329)	
Measured component: O ₃		hourly alert threshold		120	
				nmol/mol	
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	u _{r,z}	0,0013
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,600	u _{r,lv}	0,0120
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	0,400	u _{lv}	0,0768
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	0,000	u _{gp}	0,0000
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,014	u _{gt}	0,0212
6	Variations in surrounding temperature	1,0 nmol/mol/K	0,150	u _{st}	0,2700
7	Variations in electrical voltage	0,30 nmol/mol/V	-0,010	u _v	0,0147
8a	Interference H2O with 21 nmol/mol	10 nmol/mol	0,000	u _{H2O}	0,0000
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	2,147	u _{int,pos}	2,1573
8c	Interference Xylol with 0,5 µmol/mol	5,0 nmol/mol	0,397	or u _{int,neg}	
9	Averaging effect	7,0% of measured value	0,200	u _{av}	0,0192
18	Difference sample/calibration port	1,0%	0,000	u _{pac}	0,0000
23	Uncertainty test gas	3,0%	2,000	ucg	1,4400
		Combined standard uncertainty	u _c		2,0031
		Expanded uncertainty	U _c		4,0062
		Expanded uncertainty actual	U _{c,rel}		3,34
		Expanded uncertainty required	U _{req,rel}		15

Table 10 Total expanded uncertainty with the results of the laboratory test and field test according to EN 14625 (component O₃) for system 1329

Device: AR500		Serial No.		Gerät 1 (1329)		nmol/mol	
Measured component: O ₃		hourly alert threshold		120			
No.	Performance characteristic	Criterion	Result	Uncertainty		Square of uncertainty	
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	u _{r,z}	0,04	0,0013	
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,600	u _{r,iv}	not respected because, u _{r,iv} = 0,1 < u _{r,f}	-	
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	0,400	u _{lv}	0,28	0,0768	
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	0,000	u _{gp}	0,00	0,0000	
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,014	u _{gt}	0,15	0,0212	
6	Variations in surrounding temperature	1,0 nmol/mol/K	0,150	u _{st}	0,52	0,2700	
7	Variations in electrical voltage	0,30 nmol/mol/V	-0,010	u _v	-0,12	0,0147	
8a	Interference H2O with 21 nmol/mol	10 nmol/mol	0,000	u _{H2O}	0,00	0,0000	
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	2,147	u _{int,pos}	1,47	2,1573	
8c	Interference Xylol with 0,5 µmol/mol	5,0 nmol/mol	0,397	or u _{int,neg}			
9	Averaging effect	7,0% of measured value	0,200	u _{av}	0,14	0,0192	
10	Reproducibility standard deviation in field	5,0% of average of 3 month	2,410	u _{r,f}	2,89	8,3637	
11	Long term drift at zero	5,0 nmol/mol	1,460	u _{d,z}	0,84	0,7105	
12	Long term drift at span level	5,0% of range	-2,450	u _{d,lv}	-1,70	2,8812	
18	Difference sample/calibration port	1,0%	0,000	uDsc	0,00	0,0000	
23	Uncertainty test gas	3,0%	2,000	ucg	1,20	1,4400	
			Combined standard uncertainty		u _c	3,9945	nmol/mol
			Expanded uncertainty		U _c	7,9890	nmol/mol
			Expanded uncertainty actual		U _{c,rel}	6,66	%
			Expanded uncertainty required		U _{req,rel}	15	%

Table 11: Total expanded uncertainty with the results of the laboratory test according to EN 14625 (component O₃) for system 1330

Device: AR500		Serial No. Gerät 2 (1330)		120		nmol/mol	
Measured component: O ₃		hourly alert threshold					
No.	Performance characteristic	Criterion	Result	Uncertainty	Square of uncertainty		
1	Repeatability standard deviation at zero	1,0 nmol/mol	0,200	u _{r,z} 0,04	0,0013		
2	Repeatability standard deviation at ct	3,0 nmol/mol	0,400	u _{r,lv} 0,07	0,0053		
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	-0,300	u _{lv} -0,21	0,0432		
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	0,000	u _{gp} 0,00	0,0000		
5	Variations in sample gas temperature	1,0 nmol/mol/K	0,007	u _{gt} 0,07	0,0053		
6	Variations in surrounding temperature	1,0 nmol/mol/K	-0,120	u _{st} -0,42	0,1728		
7	Variations in electrical voltage	0,30 nmol/mol/V	0,010	u _v 0,12	0,0147		
8a	Interference H2O with 21 nmol/mol	10 nmol/mol	0,000	u _{H2O} 0,00	0,0000		
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	0,396	u _{int,pos} 1,72	2,9416		
8c	Interference Xylol with 0,5 µmol/mol	5,0 nmol/mol	2,574	or u _{int,neg}			
9	Averaging effect	7,0% of measured value	-0,900	u _{av} -0,62	0,3888		
18	Difference sample/calibration port	1,0%	0,000	u _{disc} 0,00	0,0000		
23	Uncertainty test gas	3,0%	2,000	0 1,20	1,4400		
		Combined standard uncertainty		u _c	2,2390	nmol/mol	
		Expanded uncertainty		U _c	4,4780	nmol/mol	
		Expanded uncertainty actual		U _{c,rel}	3,73	%	
		Expanded uncertainty required		U _{req,rel}	15	%	

Table 12

Total expanded uncertainty with the results of the laboratory test and field test according to EN 14625 (component O₃) for system 1330

Device: AR500		Serial No.	Gerät 2 (1330)	
Measured component: O ₃		hourly alert threshold	120 nmol/mol	
No.	Performance characteristic	Criterion	Result	Square of uncertainty
1	Repeatability standard deviation at zero	1,0 nmol/mol	u _{r,z}	0,0013
2	Repeatability standard deviation at ct	3,0 nmol/mol	u _{r,iv}	-
3	"lack of fit" at the hourly alert threshold value	4,0% of measured value	u _{lv}	0,0432
4	Variations in sample gas pressure	2,0 nmol/mol/kPa	u _{gp}	0,0000
5	Variations in sample gas temperature	1,0 nmol/mol/K	u _{gt}	0,0053
6	Variations in surrounding temperature	1,0 nmol/mol/K	u _{st}	0,1728
7	Variations in electrical voltage	0,30 nmol/mol/V	u _v	0,0147
8a	Interference H2O with 21 mmol/mol	10 nmol/mol	u _{H2O}	0,0000
8b	Interference Toluol with 0,5 µmol/mol	5,0 nmol/mol	u _{int,pos}	2,9416
8c	Interference Xylol with 0,5 µmol/mol	5,0 nmol/mol	u _{int,neg}	2,9416
9	Averaging effect	7,0% of measured value	u _{av}	0,3888
10	Reproducibility standard deviation in field	5,0% of average of 3 month	u _{rf}	8,3637
11	Long term drift at zero	5,0 nmol/mol	u _{d,l,z}	1,1285
12	Long term drift at span level	5,0% of range	u _{d,lv}	4,0368
18	Difference sample/calibration port	1,0%	u _{Dsc}	0,0000
23	Uncertainty test gas	3,0%	0	1,4400
		Combined standard uncertainty	u _c	4,3054
		Expanded uncertainty	U _c	8,6109
		Expanded uncertainty actual	U _{c,rel}	7,18
		Expanded uncertainty required	U _{req,rel}	15
				%